

Application No.: 09/447,227
Filing Date. November 22, 1999

SUMMARY OF INTERVIEW

Attendees, Date and Type of Interview

Applicants thank Examiner Robert Nasser for the telephonic interview with Applicants' representative, Laura Johnson, on May 26, 2010 to discuss the pending application.

Exhibits and/or Demonstrations

None.

Identification of Claims Discussed

Claims 88, 96, and proposed new claims.

Identification of Prior Art Discussed

None.

Proposed Amendments, Principal Arguments and Other Matters, Results of Interviews

Applicants' representative proposed amending the claims to replace the term "sensing mechanism" with the term "sensor." Applicants' representative also proposed a new independent claim which recites that the device is capable of an accuracy level in which at least 95% of glucose concentration values measured by the device are within 25% of corresponding values determined by analysis of blood over a time period of at least 2 days. The Examiner agreed that the proposed amendments and the new claims appeared to be supported in the specification as filed, but further consideration would be required.